

#### 305.4 - Flooring Radiant Panel (sheet form)

This SRM consists of three sheets of kraft paperboard. It is for checking the operation of flooring radiant panel test apparatus used to measure critical radiant flux as per ASTM E 648.

PLEASE NOTE: The tables are presented to facilitate comparisons among a family of materials to help customers select the best SRM for their needs. For specific values and uncertainties, the certificate is the only official source.

SRM	Description	Unit Size	Critical Radiant Flux
1012	Flooring Radiant Panel	set (3)	0.36 W/cm <sup>2</sup>

- Certified values are normal font
- Reference values are italicized
- Values in parentheses are for information only